

BASIC Mobile X-Ray System 60mA/100mA





BASIC MOBILE X-RAY SYSTEM SPECIFICATION

MODEL NAME	GXL-D60	GXL-D100
POWER SUPPLY REQUIREMENT	1 Phase	1 Phase
INPUT RATING	230 VAC/ 15 A / 50 Hz/ ±10%	230 VAC/ 15 A / 50 Hz/ ±10%
LINE RESISTANCE	0.6 ohms	0.4 ohms
OUTPUT RATING	4.8 kW	8 kW
CONTROL PANEL	Microprocessor Controlled Feather Touch Operated	Microprocessor Controlled Feather Touch Operated
RAD kVp RANGE	40 - 100 kVp	40 - 100 kVp
RAD mAS RANGE	2 - 250 mAS	2 - 250 mAS
RAD mA RANGE	25, 40, 50, 60	25, 40, 60, 80, 100
EXPOSURE TIME	0.04 - 5.0 Sec.	0.02 - 5.0 Sec.
X-RAY TUBE	STATIONARY ANODE	STATIONARY ANODE
TUBEHEAD TYPE	Monoblock	Monoblock
RECTIFICATION	Full Wave Rectification	Full Wave Rectification
X-RAY TUBE FOCAL SPOT	2.8 mm ²	2.8 mm ²
MOBILE TUBE STAND	GXL-PBS	GXL-PBS
* OPTIONAL BUCKY TABLE	Moving Bucky With Grid	Moving Bucky With Grid
	GXL-Horizontal Table	GXL-Horizontal Table
	GXL-2Way Floating Table	GXL-2Way Floating Table
	GXL-Folding Table	GXL-Folding Table

Optional Bucky Table

